



California Medical Device Recall Information



Recall Name

Baxter Healthcare Recalls Buretrol Solution Sets Due to a Possible Malfunction

Recall Date	Product Description	Recalling Firm	Recall Reason
9/07/12	<ul style="list-style-type: none">Interlink System Buretrol Solution SetClearlink System Buretrol Solution Set <p>The products are non-reusable, disposable medical devices for the administration of fluids directly into blood vessels.</p>	Baxter Healthcare, Corp. Deerfield, IL	<i>Ball-valve feature may not function as expected</i>
Recall Class	Product Identification	Distribution	Affected Dates
I	<p>Product Codes affected:</p> <ul style="list-style-type: none">2C75192H75192C88192H8819 <p>All Lots are Recalled</p>	CA , nationwide	Distributed from May 1, 2003 thru August 16, 2012

FOR ADDITIONAL INFORMATION, PLEASE VISIT:

<http://www.fda.gov/MedicalDevices/Safety/ListofRecalls/ucm327524.htm>